
Midmark M11 - v1.0.0
Total Cycles: 4016

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2023 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 79.0 F 0.1
4:00 85.4 F 1.0
6:00 119.4 F 3.2
8:00 145.7 F 5.7
10:00 168.4 F 8.4
12:00 184.0 F 11.6
14:00 204.6 F 2.2
16:00 211.6 F 5.4
18:00 220.3 F 8.7
20:00 230.2 F 12.3
22:00 231.1 F 7.5
24:00 242.0 F 12.4
26:00 239.8 F 10.2
28:00 251.7 F 15.9
30:00 262.0 F 21.8
32:00 271.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.6 F 28.8
2:00 271.6 F 28.1
3:00 272.0 F 28.7
4:00 272.1 F 28.6

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.9 F 0.0
5:00 194.8 F 0.0
10:00 190.8 F 0.0
15:00 188.2 F 0.0
20:00 201.0 F 0.0
25:00 206.2 F 0.0
30:00 208.2 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:38
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:11:09
CYCLE COMPLETE

11/14/2023 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____