
Midmark M11 - v1.0.5
Total Cycles: 1457

Name: RITTER2
Sterilizer ID: V2313228
09/06/2022 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.9 F 0.0
2:00 183.9 F 2.3
4:00 196.0 F 7.0
6:00 215.0 F 12.6
8:00 226.6 F 9.0
10:00 235.1 F 9.8
12:00 243.6 F 12.9
14:00 243.2 F 12.0
16:00 256.0 F 18.8
18:00 267.2 F 25.7
20:00 273.4 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.7
1:00 272.4 F 29.7
2:00 272.5 F 29.7
3:00 272.3 F 29.9
4:00 272.5 F 30.1

Min 272.0 F 29.6
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 217.5 F 0.0
10:00 236.5 F 0.0
15:00 237.0 F 0.0
20:00 223.2 F 0.0
25:00 224.7 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 2:26
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

09/06/2022 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____