

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.2uS / 0.7ppm  
CYCLE NUMBER 000620  
11:24 01/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:09  
PRESSURIZING 3:16  
STERILIZING 4:36  
136.0 C 227kPa 4:36  
136.3 C 231kPa 5:06  
136.4 C 231kPa 5:36  
136.5 C 232kPa 6:06  
136.4 C 232kPa 6:36  
136.3 C 231kPa 7:06  
136.4 C 231kPa 7:36  
136.3 C 231kPa 8:06  
136.3 C 231kPa 8:06  
VENTING 8:06  
AIR DRYING 9:01  
CYCLE COMPLETE 24:01  
-----

Digital Signature #  
9A494C808F535ADB  
-----