
Midmark M11 - v1.0.5
Total Cycles: 4299

Name: RITTER2
Sterilizer ID: V2313228
01/27/2025 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.4 F 0.0
2:00 136.2 F 0.9
4:00 144.0 F 3.4
6:00 151.2 F 6.9
8:00 178.5 F 10.8
10:00 197.1 F 4.8
12:00 212.3 F 7.5
14:00 226.8 F 11.8
16:00 228.9 F 8.6
18:00 239.8 F 12.4
20:00 241.2 F 12.4
22:00 252.7 F 18.2
24:00 263.4 F 24.5
26:00 272.6 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.9
1:00 272.6 F 30.8
2:00 272.5 F 30.4
3:00 271.6 F 29.9
4:00 271.4 F 30.2

Min 271.2 F 29.8
Max 273.2 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 203.4 F 0.0
10:00 208.5 F 0.0
15:00 198.3 F 0.0
20:00 196.9 F 0.0
25:00 245.5 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:19
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:05:12
CYCLE COMPLETE

01/27/2025 02:57 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____