

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1757

Name: RITTER 1  
Sterilizer ID: V2178198  
03/28/2022 04:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.8 F 0.0  
2:00 176.2 F 3.4  
4:00 189.6 F 8.0  
6:00 210.5 F 11.4  
8:00 225.9 F 8.9  
10:00 237.1 F 9.3  
12:00 242.1 F 12.5  
14:00 241.4 F 11.2  
16:00 255.2 F 18.2  
18:00 267.1 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.4  
1:00 272.1 F 29.1  
2:00 272.1 F 28.8  
3:00 271.5 F 28.6  
4:00 272.6 F 29.2  
  
Min 271.3 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.6 F 0.0  
5:00 201.3 F 0.0  
10:00 209.9 F 0.0  
15:00 213.0 F 0.0  
20:00 244.0 F 0.0  
25:00 230.6 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 19:16  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 00:57:57  
CYCLE COMPLETE

03/28/2022 05:33 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_