

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5327

Name: RITTER 1  
Sterilizer ID: V2178198  
02/10/2025 11:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.2 F 0.0  
2:00 160.6 F 2.3  
4:00 166.1 F 5.4  
6:00 182.5 F 8.6  
8:00 208.8 F 12.1  
10:00 213.9 F 3.9  
12:00 220.6 F 7.7  
14:00 231.6 F 11.7  
16:00 233.4 F 7.5  
18:00 244.3 F 12.5  
20:00 241.1 F 10.5  
22:00 252.2 F 16.0  
24:00 261.8 F 21.5  
26:00 270.5 F 27.2  
28:00 271.5 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 271.9 F 28.7  
2:00 272.4 F 29.1  
3:00 272.7 F 29.2  
4:00 272.7 F 29.1

Min 271.3 F 28.3  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
5:00 230.3 F 0.0  
10:00 216.8 F 0.0  
15:00 231.4 F 0.0  
20:00 232.8 F 0.0  
25:00 215.8 F 0.0  
30:00 214.0 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 28:04  
STERILIZING: 4:00  
VENTING: 4:05  
DRYING: 30:00  
TOTAL CYCLE: 01:07:06  
CYCLE COMPLETE

02/10/2025 12:39 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_