
Midmark M11 - v1.0.5
Total Cycles: 2461

Name: RITTER2
Sterilizer ID: V2313228
06/05/2023 07:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 76.2 F 0.0
4:00 80.6 F 0.0
6:00 103.9 F 0.0
8:00 139.9 F 1.8
10:00 168.4 F 4.9
12:00 196.0 F 8.4
14:00 207.7 F 12.7
16:00 216.4 F 6.1
18:00 228.4 F 10.1
20:00 233.7 F 9.3
22:00 239.9 F 11.8
24:00 237.9 F 9.9
26:00 250.7 F 16.1
28:00 261.8 F 22.4
30:00 271.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.0 F 29.6
2:00 271.9 F 29.0
3:00 272.0 F 29.6
4:00 272.1 F 29.2

Min 271.2 F 28.8
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 207.0 F 0.0
10:00 216.9 F 0.0
15:00 230.0 F 0.0
20:00 227.7 F 0.0
25:00 229.3 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:25
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:23
CYCLE COMPLETE

06/05/2023 08:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____