

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002645  
8:12 02/04/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:28  
STERILIZING 3:29  
136.4 C 223kPa 3:29  
136.2 C 222kPa 3:59  
136.2 C 223kPa 4:29  
136.3 C 222kPa 4:59  
136.2 C 223kPa 5:29  
136.4 C 224kPa 5:59  
136.3 C 224kPa 6:29  
136.0 C 222kPa 6:59  
136.0 C 222kPa 6:59  
VENTING 6:59  
AIR DRYING 7:49  
CYCLE COMPLETE 18:17  
-----

Digital Signature #  
6B94DE421EB87746  
-----