
Midmark M11 - v1.0.5
Total Cycles: 3374

Name: RITTER2
Sterilizer ID: V2313228
02/09/2024 03:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.4 F 0.0
2:00 95.0 F 0.0
4:00 113.9 F 0.5
6:00 152.6 F 2.8
8:00 176.7 F 5.6
10:00 187.5 F 9.0
12:00 204.5 F 12.4
14:00 216.7 F 5.7
16:00 226.6 F 9.5
18:00 232.1 F 10.8
20:00 238.1 F 11.1
22:00 236.2 F 9.5
24:00 249.7 F 15.7
26:00 260.9 F 22.1
28:00 271.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.4 F 30.0
2:00 271.8 F 29.3
3:00 271.7 F 29.8
4:00 272.3 F 29.7

Min 271.1 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 231.4 F 0.0
10:00 212.0 F 0.0
15:00 194.6 F 0.0
20:00 188.5 F 0.0
25:00 184.6 F 0.0
30:00 181.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:31
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:07:13
CYCLE COMPLETE

02/09/2024 04:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____