
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005339
12:29 10/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 1:59
STERILIZING 2:56
136.4 C 220kPa 2:56
136.4 C 224kPa 3:26
136.1 C 217kPa 3:56
136.4 C 223kPa 4:26
136.4 C 222kPa 4:56
136.0 C 220kPa 5:26
136.4 C 224kPa 5:56
136.4 C 222kPa 6:26
136.4 C 222kPa 6:26
VENTING 6:26
AIR DRYING 7:18
CYCLE COMPLETE 10:10

Digital Signature #
CCA6EFB3FEAF7068
