

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4687

Name: RITTER 1  
Sterilizer ID: V2178198  
06/28/2024 09:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.2 F 0.0  
2:00 136.4 F 1.5  
4:00 146.6 F 4.2  
6:00 180.5 F 7.3  
8:00 203.5 F 11.5  
10:00 217.9 F 3.4  
12:00 228.3 F 7.4  
14:00 232.4 F 11.6  
16:00 232.1 F 7.3  
18:00 243.2 F 12.6  
20:00 240.7 F 10.5  
22:00 252.4 F 16.3  
24:00 262.9 F 22.4  
26:00 272.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 28.4  
1:00 272.5 F 28.9  
2:00 271.6 F 28.7  
3:00 272.2 F 28.5  
4:00 272.4 F 29.0

Min 271.3 F 28.3  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 234.6 F 0.0  
10:00 235.0 F 0.0  
15:00 229.5 F 0.0  
20:00 223.1 F 0.0  
25:00 227.1 F 0.0  
30:00 230.2 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 26:59  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:05:35  
CYCLE COMPLETE

06/28/2024 10:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_