

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007

Unit # : 000

WATER QUALITY

0.0uS / 0.0ppm

CYCLE NUMBER 002042

9:33 16/09/2024

WRAPPED

135 C for 10 min.

CYCLE START 0:00

CONDITIONING 2:49

PRESSURIZING 4:56

STERILIZING 5:50

136.4 C 226kPa 5:50

136.2 C 224kPa 6:20

136.3 C 225kPa 6:50

136.4 C 226kPa 7:20

136.4 C 224kPa 7:50

136.2 C 224kPa 8:20

136.4 C 226kPa 8:50

136.4 C 225kPa 9:20

136.4 C 226kPa 9:50

136.2 C 225kPa 10:20

136.1 C 224kPa 10:50

136.3 C 225kPa 11:20

136.4 C 225kPa 11:50

136.3 C 224kPa 12:20

136.3 C 224kPa 12:50

136.4 C 226kPa 13:20

136.4 C 221kPa 13:50

136.4 C 226kPa 14:20

136.1 C 222kPa 14:50

136.2 C 224kPa 15:20

136.3 C 225kPa 15:50

136.3 C 225kPa 15:50

VENTING 15:50

AIR DRYING 16:44

CYCLE COMPLETE 16:57  
-----

Digital Signature #  
4E2B9469FF08C875  
-----