
Midmark M11 - v1.0.0
Total Cycles: 3327

Name: RITTER 1
Sterilizer ID: V2178198
06/02/2023 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.3 F 0.0
2:00 154.8 F 0.0
4:00 165.2 F 0.2
6:00 190.2 F 3.8
8:00 207.4 F 7.8
10:00 219.1 F 12.0
12:00 225.1 F 6.7
14:00 234.5 F 11.4
16:00 234.0 F 8.4
18:00 242.7 F 12.0
20:00 245.0 F 12.7
22:00 256.5 F 18.8
24:00 266.8 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.5 F 28.9
2:00 271.7 F 28.3
3:00 271.4 F 28.5
4:00 272.4 F 28.9

Min 271.1 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 208.9 F 0.0
10:00 218.4 F 0.0
15:00 223.8 F 0.0
20:00 234.2 F 0.0
25:00 225.8 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 25:37
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:04:05
CYCLE COMPLETE

06/02/2023 12:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____