

-----  
Midmark M11 - v1.0.0  
Total Cycles: 990

Name: RITTER 1  
Sterilizer ID: V2178198  
07/08/2021 05:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.1 F 0.0  
2:00 194.7 F 4.8  
4:00 207.4 F 10.3  
6:00 221.3 F 5.4  
8:00 238.4 F 11.4  
10:00 235.7 F 9.0  
12:00 238.2 F 9.0  
14:00 249.8 F 15.0  
16:00 262.9 F 22.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.1 F 28.4  
2:00 272.7 F 28.9  
3:00 271.7 F 28.1  
4:00 272.5 F 28.6  
  
Min 271.5 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.5 F 0.0  
5:00 231.7 F 0.0  
10:00 225.4 F 0.0  
15:00 224.2 F 0.0  
20:00 224.6 F 0.0  
25:00 224.7 F 0.0  
30:00 222.5 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 17:60  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 00:56:17  
CYCLE COMPLETE

07/12/2021 06:21 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01