
Midmark M11 - v1.0.0
Total Cycles: 3795

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2023 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.4 F 0.0
2:00 135.9 F 0.0
4:00 159.7 F 0.8
6:00 185.9 F 3.5
8:00 204.6 F 6.9
10:00 223.5 F 10.9
12:00 224.5 F 5.9
14:00 234.0 F 9.5
16:00 239.7 F 10.9
18:00 241.9 F 11.5
20:00 238.9 F 9.8
22:00 252.0 F 16.1
24:00 263.1 F 22.5
26:00 272.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.2 F 28.5
2:00 272.1 F 28.7
3:00 271.8 F 28.2
4:00 272.4 F 28.9

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.6 F 0.0
5:00 216.2 F 0.0
10:00 202.4 F 0.0
15:00 210.1 F 0.0
20:00 210.8 F 0.0
25:00 209.4 F 0.0
30:00 214.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:14
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:04:43
CYCLE COMPLETE

09/19/2023 01:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____