
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000442
13:12 28/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:43
STERILIZING 3:52
136.4 C 233kPa 3:52
136.2 C 231kPa 4:22
136.4 C 233kPa 4:52
136.4 C 233kPa 5:22
136.2 C 231kPa 5:52
136.4 C 233kPa 6:22
136.4 C 233kPa 6:52
136.4 C 233kPa 7:22
136.4 C 233kPa 7:22
VENTING 7:22
AIR DRYING 8:14
CYCLE COMPLETE 8:36

Digital Signature #
73F834F2B8585405
