

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2825

Name: RITTER 1  
Sterilizer ID: V2178198  
01/30/2023 12:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.9 F 0.0  
2:00 156.7 F 3.4  
4:00 177.3 F 7.3  
6:00 210.5 F 11.8  
8:00 215.4 F 5.4  
10:00 228.3 F 10.4  
12:00 229.5 F 7.0  
14:00 241.0 F 12.5  
16:00 240.9 F 10.7  
18:00 253.9 F 17.3  
20:00 265.1 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.6 F 29.0  
2:00 272.0 F 29.1  
3:00 272.2 F 28.9  
4:00 271.9 F 29.1  
  
Min 271.6 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.6 F 0.0  
5:00 215.2 F 0.0  
10:00 230.4 F 0.0  
15:00 236.1 F 0.0  
20:00 230.4 F 0.0  
25:00 225.1 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 21:46  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:04  
CYCLE COMPLETE

01/30/2023 01:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_