
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001722
10:54 27/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:52
STERILIZING 4:00
136.4 C 224kPa 4:00
136.3 C 225kPa 4:31
136.4 C 225kPa 5:01
136.4 C 225kPa 5:31
136.4 C 222kPa 6:01
136.3 C 224kPa 6:31
136.3 C 224kPa 7:01
136.4 C 225kPa 7:31
136.4 C 225kPa 7:31
VENTING 7:31
AIR DRYING 8:22
CYCLE COMPLETE 8:37

Digital Signature #
83A7DA9B6DC5D200
