
Midmark M11 - v1.0.0
Total Cycles: 710

Name: MIDMARK
Sterilizer ID: V2178198
03/01/2021 07:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 81.5 F 0.0
4:00 140.0 F 0.1
6:00 190.0 F 0.1
8:00 209.5 F 0.2
10:00 215.5 F 1.5
12:00 225.8 F 5.1
14:00 226.0 F 8.2
16:00 231.6 F 11.9
18:00 233.4 F 7.7
20:00 245.4 F 13.0
22:00 240.8 F 10.7
24:00 236.5 F 8.9
26:00 251.7 F 16.0
28:00 263.9 F 23.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.3 F 28.4
2:00 272.0 F 28.9
3:00 271.4 F 28.0
4:00 272.4 F 28.8

Min 271.0 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 220.7 F 0.0
10:00 238.5 F 0.0
15:00 200.4 F 0.0
20:00 158.9 F 0.0
25:00 149.9 F 0.0
30:00 181.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:58
STERILIZING: 4:00
VENTING: 2:53
DRYING: 30:00
TOTAL CYCLE: 01:07:39
CYCLE COMPLETE

03/01/2021 08:40 AM
ACCEPTED: X
REJECTED:
Operator: U01