

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4637

Name: RITTER 1  
Sterilizer ID: V2178198  
06/13/2024 01:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.3 F 0.0  
2:00 121.0 F 1.3  
4:00 137.3 F 3.7  
6:00 163.0 F 6.9  
8:00 192.9 F 10.8  
10:00 214.1 F 3.4  
12:00 218.6 F 7.8  
14:00 230.2 F 12.2  
16:00 232.9 F 8.3  
18:00 243.3 F 12.4  
20:00 243.4 F 12.1  
22:00 255.1 F 18.1  
24:00 265.6 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 271.9 F 28.6  
2:00 272.5 F 29.0  
3:00 271.5 F 29.0  
4:00 271.3 F 28.5

Min 271.0 F 28.4  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
5:00 214.8 F 0.0  
10:00 222.2 F 0.0  
15:00 234.2 F 0.0  
20:00 224.0 F 0.0  
25:00 208.3 F 0.0  
30:00 196.6 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:44  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:04:14  
CYCLE COMPLETE

06/13/2024 02:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_