

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3647

Name: RITTER 1  
Sterilizer ID: V2178198  
08/14/2023 04:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.4 F 0.0  
2:00 176.9 F 3.3  
4:00 203.2 F 7.9  
6:00 218.8 F 12.5  
8:00 226.3 F 7.2  
10:00 235.7 F 12.1  
12:00 236.6 F 9.4  
14:00 238.8 F 9.7  
16:00 248.3 F 14.3  
18:00 260.0 F 20.8  
20:00 270.7 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.1 F 28.4  
2:00 272.5 F 28.8  
3:00 271.9 F 28.8  
4:00 271.8 F 28.3

Min 271.3 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.1 F 0.0  
5:00 203.9 F 0.0  
10:00 209.6 F 0.0  
15:00 222.7 F 0.0  
20:00 226.9 F 0.0  
25:00 229.4 F 0.0  
30:00 230.1 F 0.0  
Drying Complete

FILLING: 2:07  
HEATING: 20:43  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:00:28  
CYCLE COMPLETE

08/14/2023 05:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_