

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.2uS / 0.7ppm  
CYCLE NUMBER 001107  
11:51 03/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:07  
PRESSURIZING 3:15  
STERILIZING 4:48  
136.4 C 225kPa 4:48  
136.2 C 224kPa 5:18  
136.3 C 225kPa 5:48  
136.4 C 226kPa 6:18  
136.4 C 226kPa 6:48  
136.4 C 226kPa 7:18  
136.4 C 225kPa 7:48  
136.4 C 227kPa 8:18  
136.4 C 227kPa 8:18  
VENTING 8:18  
AIR DRYING 9:17  
CYCLE COMPLETE 24:17  
-----

Digital Signature #  
AA26FC50862EDDC4  
-----