
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004208
11:52 16/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:23
STERILIZING 3:16
136.0 C 220kPa 3:16
136.3 C 222kPa 3:46
136.4 C 223kPa 4:16
136.2 C 223kPa 4:46
136.4 C 225kPa 5:16
136.1 C 221kPa 5:46
136.1 C 222kPa 6:16
136.1 C 222kPa 6:46
136.1 C 222kPa 6:46
VENTING 6:46
AIR DRYING 7:38
CYCLE COMPLETE 7:46

Digital Signature #
18BA9CBA7D785743
