
Midmark M11 - v1.0.0
Total Cycles: 3515

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2023 08:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 72.5 F 0.0
4:00 87.1 F 0.0
6:00 108.6 F 0.0
8:00 141.9 F 0.9
10:00 168.8 F 2.8
12:00 186.8 F 5.6
14:00 199.8 F 8.8
16:00 213.3 F 12.2
18:00 220.6 F 5.4
20:00 230.9 F 9.4
22:00 239.5 F 12.2
24:00 238.8 F 10.6
26:00 235.1 F 7.9
28:00 248.7 F 14.7
30:00 259.8 F 20.8
32:00 269.9 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.1
2:00 271.5 F 28.2
3:00 272.8 F 29.2
4:00 271.8 F 28.5

Min 271.5 F 28.2
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 198.8 F 0.0
10:00 217.2 F 0.0
15:00 224.3 F 0.0
20:00 226.4 F 0.0
25:00 229.7 F 0.0
30:00 228.2 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:59
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:11:02
CYCLE COMPLETE

07/12/2023 09:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____