
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001532
15:07 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:11
STERILIZING 4:14
136.4 C 225kPa 4:14
136.2 C 225kPa 4:44
135.9 C 221kPa 5:14
136.4 C 227kPa 5:44
136.4 C 222kPa 6:14
136.4 C 226kPa 6:44
136.4 C 225kPa 7:14
136.4 C 227kPa 7:44
136.4 C 227kPa 7:44
VENTING 7:44
AIR DRYING 8:39
CYCLE COMPLETE 8:51

Digital Signature #
CCC20B63FC92C195
