
Midmark M11 - v1.0.5
Total Cycles: 365

Name: RITTER2
Sterilizer ID: V2313228
09/15/2021 08:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.1 F 0.0
2:00 99.9 F 0.5
4:00 115.3 F 2.3
6:00 142.1 F 5.0
8:00 170.2 F 8.9
10:00 192.3 F 13.1
12:00 222.5 F 6.2
14:00 229.1 F 10.3
16:00 232.5 F 8.6
18:00 240.3 F 11.7
20:00 237.8 F 10.0
22:00 250.1 F 16.0
24:00 261.2 F 22.2
26:00 271.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.9 F 30.0
2:00 271.4 F 29.2
3:00 272.0 F 30.0
4:00 272.7 F 29.7

Min 271.2 F 29.1
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 236.2 F 0.0
10:00 192.0 F 0.0
15:00 178.6 F 0.0
20:00 177.6 F 0.0
25:00 210.0 F 0.0
30:00 196.2 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:46
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:05:01
CYCLE COMPLETE

09/15/2021 09:36 AM
ACCEPTED: X
REJECTED:
Operator: U01