

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2678

Name: RITTER2  
Sterilizer ID: V2313228  
07/21/2023 08:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.5 F 0.0  
2:00 74.3 F 0.0  
4:00 96.9 F 0.0  
6:00 135.4 F 0.0  
8:00 143.8 F 1.4  
10:00 167.5 F 4.3  
12:00 195.6 F 7.1  
14:00 203.8 F 11.2  
16:00 218.2 F 4.9  
18:00 226.8 F 8.7  
20:00 236.8 F 13.0  
22:00 235.6 F 9.8  
24:00 237.6 F 9.5  
26:00 245.7 F 13.7  
28:00 256.7 F 19.6  
30:00 266.7 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.8 F 29.8  
2:00 271.6 F 29.0  
3:00 272.2 F 29.9  
4:00 272.3 F 29.5

Min 271.3 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 205.1 F 0.0  
10:00 210.5 F 0.0  
15:00 213.5 F 0.0  
20:00 211.0 F 0.0  
25:00 211.1 F 0.0  
30:00 212.7 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:36  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:09:33  
CYCLE COMPLETE

07/21/2023 09:55 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_