
Midmark M11 - v1.0.5
Total Cycles: 3556

Name: RITTER2
Sterilizer ID: V2313228
04/24/2024 04:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.7 F 0.0
2:00 119.6 F 1.4
4:00 133.6 F 4.0
6:00 163.9 F 7.4
8:00 181.9 F 11.6
10:00 210.6 F 4.9
12:00 222.4 F 9.2
14:00 234.5 F 12.6
16:00 234.4 F 10.4
18:00 234.8 F 8.6
20:00 246.7 F 14.3
22:00 257.9 F 20.4
24:00 268.2 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.7
1:00 272.6 F 30.1
2:00 271.8 F 30.3
3:00 271.5 F 29.6
4:00 272.1 F 30.1

Min 271.2 F 29.6
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 227.1 F 0.0
10:00 236.9 F 0.0
15:00 226.4 F 0.0
20:00 228.9 F 0.0
25:00 229.4 F 0.0
30:00 231.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 25:56
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:04:36
CYCLE COMPLETE

04/24/2024 05:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____