

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002485  
9:00 07/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:05  
PRESSURIZING 3:10  
STERILIZING 4:16  
136.5 C 224kPa 4:16  
136.2 C 223kPa 4:46  
136.2 C 222kPa 5:16  
136.4 C 225kPa 5:46  
135.9 C 221kPa 6:16  
136.1 C 221kPa 6:46  
136.0 C 220kPa 7:16  
136.3 C 223kPa 7:46  
136.3 C 223kPa 7:46  
VENTING 7:46  
AIR DRYING 8:35  
CYCLE COMPLETE 31:06  
-----

Digital Signature #  
C3CB301F237A82CF  
-----