

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3065

Name: RITTER 1  
Sterilizer ID: V2178198  
04/03/2023 08:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.5 F 0.0  
2:00 80.4 F 0.0  
4:00 118.4 F 0.0  
6:00 136.8 F 0.5  
8:00 163.5 F 2.5  
10:00 179.9 F 5.9  
12:00 196.2 F 9.8  
14:00 209.0 F 8.9  
16:00 222.6 F 6.5  
18:00 233.5 F 10.7  
20:00 232.6 F 7.5  
22:00 244.1 F 12.8  
24:00 242.2 F 11.3  
26:00 255.4 F 18.0  
28:00 266.8 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 271.6 F 27.9  
2:00 271.9 F 28.1  
3:00 272.3 F 28.4  
4:00 272.6 F 28.7

Min 271.4 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
5:00 201.4 F 0.0  
10:00 211.8 F 0.0  
15:00 215.3 F 0.0  
20:00 217.8 F 0.0  
25:00 217.5 F 0.0  
30:00 217.3 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 29:23  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 01:07:11  
CYCLE COMPLETE

04/03/2023 09:24 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_