
Midmark M11 - v1.0.0
Total Cycles: 2632

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2022 05:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.4 F 0.0
2:00 152.4 F 2.3
4:00 178.7 F 6.1
6:00 212.1 F 10.3
8:00 220.0 F 4.2
10:00 227.5 F 8.5
12:00 238.5 F 13.4
14:00 238.1 F 10.4
16:00 236.3 F 8.1
18:00 249.0 F 14.9
20:00 260.8 F 21.5
22:00 271.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.2 F 29.2
2:00 271.4 F 28.4
3:00 272.1 F 29.4
4:00 271.8 F 28.9

Min 271.2 F 28.4
Max 272.2 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 210.4 F 0.0
10:00 225.4 F 0.0
15:00 229.5 F 0.0
20:00 230.3 F 0.0
25:00 229.3 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

11/30/2022 06:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____