
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002726
10:56 24/11/2021
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 4:45
STERILIZING 5:34
136.0 C 220kPa 5:34
136.4 C 223kPa 6:04
136.4 C 223kPa 6:34
136.2 C 222kPa 7:04
136.4 C 224kPa 7:34
136.4 C 223kPa 8:04
136.4 C 225kPa 8:34
136.4 C 223kPa 9:04
136.3 C 222kPa 9:34
136.4 C 221kPa 10:04
136.4 C 224kPa 10:34
136.4 C 223kPa 11:04
136.3 C 221kPa 11:34
136.3 C 223kPa 12:04
136.4 C 222kPa 12:34
136.1 C 220kPa 13:04
136.4 C 224kPa 13:34
136.4 C 222kPa 14:04
136.2 C 218kPa 14:34
136.4 C 224kPa 15:04
136.1 C 218kPa 15:34
136.1 C 218kPa 15:34
VENTING 15:34
AIR DRYING 16:31
CYCLE COMPLETE 16:32

Digital Signature #

D25EB630A3B7200D
