
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001038
14:31 13/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:26
STERILIZING 3:34
136.5 C 224kPa 3:34
136.2 C 224kPa 4:04
136.4 C 226kPa 4:34
136.3 C 224kPa 5:04
136.2 C 225kPa 5:34
136.4 C 226kPa 6:04
136.1 C 223kPa 6:34
136.4 C 225kPa 7:04
136.4 C 225kPa 7:04
VENTING 7:04
AIR DRYING 7:59
CYCLE COMPLETE 21:42

Digital Signature #
B5E14A6E63EE277B
