
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002002
11:14 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:47
STERILIZING 3:54
136.0 C 220kPa 3:54
136.4 C 226kPa 4:24
136.4 C 226kPa 4:54
136.4 C 226kPa 5:24
136.3 C 226kPa 5:54
136.1 C 224kPa 6:24
136.4 C 223kPa 6:54
136.3 C 225kPa 7:24
136.3 C 225kPa 7:24
VENTING 7:24
AIR DRYING 8:26
CYCLE COMPLETE 9:05

Digital Signature #
4EE3991054AA4C40
