

-----  
Midmark M11 - v1.0.0  
Total Cycles: 625

Name: MIDMARK  
Sterilizer ID: V2178198  
01/25/2021 02:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.8 F 0.0  
2:00 121.8 F 0.8  
4:00 160.7 F 3.1  
6:00 179.1 F 7.4  
8:00 202.5 F 12.8  
10:00 218.4 F 6.2  
12:00 238.7 F 11.4  
14:00 235.8 F 8.4  
16:00 240.6 F 10.2  
18:00 247.0 F 13.4  
20:00 260.0 F 20.5  
22:00 271.7 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.4  
1:00 271.8 F 28.6  
2:00 272.0 F 28.8  
3:00 271.9 F 28.8  
4:00 271.6 F 28.6

Min 271.5 F 28.3  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.1 F 0.0  
5:00 220.0 F 0.0  
10:00 194.2 F 0.0  
15:00 190.6 F 0.0  
20:00 229.9 F 0.0  
25:00 233.4 F 0.0  
Dry Cycle Aborted

FILLING: 0:58  
HEATING: 23:03  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 25:51  
TOTAL CYCLE: 00:56:48

01/25/2021 03:01 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01