
Midmark M11 - v1.0.0
Total Cycles: 1302

Name: RITTER 1
Sterilizer ID: V2178198
10/26/2021 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.6 F 0.0
2:00 106.8 F 0.0
4:00 122.2 F 0.1
6:00 154.2 F 0.2
8:00 178.8 F 2.9
10:00 200.5 F 6.1
12:00 218.6 F 9.6
14:00 224.5 F 11.0
16:00 232.3 F 8.5
18:00 242.3 F 13.1
20:00 239.4 F 10.3
22:00 236.9 F 8.5
24:00 248.7 F 14.6
26:00 259.9 F 20.8
28:00 270.5 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.8 F 29.0
2:00 271.5 F 28.1
3:00 272.6 F 29.1
4:00 272.3 F 28.5

Min 271.3 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 234.0 F 0.0
10:00 224.5 F 0.0
15:00 208.6 F 0.0
20:00 196.3 F 0.0
25:00 191.2 F 0.0
30:00 189.4 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 28:53
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:07:16
CYCLE COMPLETE

10/26/2021 02:09 PM
ACCEPTED: X
REJECTED:
Operator: U01