

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2840

Name: RITTER 1  
Sterilizer ID: V2178198  
02/01/2023 10:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.7 F 0.0  
2:00 197.1 F 4.4  
4:00 197.7 F 9.2  
6:00 210.1 F 8.4  
8:00 226.2 F 8.5  
10:00 238.2 F 10.9  
12:00 239.5 F 11.2  
14:00 238.0 F 9.5  
16:00 252.0 F 16.4  
18:00 264.0 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.4  
1:00 272.1 F 29.0  
2:00 272.3 F 28.8  
3:00 271.5 F 28.4  
4:00 272.5 F 29.3  
  
Min 271.5 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 233.1 F 0.0  
10:00 230.1 F 0.0  
15:00 229.2 F 0.0  
20:00 227.0 F 0.0  
25:00 226.3 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 19:51  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:24  
CYCLE COMPLETE

02/01/2023 11:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_