
Midmark M11 - v1.0.5
Total Cycles: 92

Name: RITTER2
Sterilizer ID: V2313228
06/14/2021 07:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 73.5 F 0.1
4:00 82.8 F 0.8
6:00 147.1 F 2.7
8:00 141.9 F 5.5
10:00 169.3 F 9.0
12:00 191.7 F 12.8
14:00 219.7 F 6.1
16:00 228.1 F 9.6
18:00 236.8 F 12.7
20:00 235.9 F 9.6
22:00 237.9 F 10.4
24:00 244.2 F 12.7
26:00 255.1 F 18.4
28:00 265.2 F 24.6
30:00 270.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 271.6 F 29.5
2:00 271.4 F 29.5
3:00 271.2 F 29.5
4:00 271.4 F 29.3

Min 271.2 F 29.2
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 231.5 F 0.0
10:00 211.4 F 0.0
15:00 203.4 F 0.0
20:00 192.5 F 0.0
25:00 190.1 F 0.0
Dry Cycle Aborted

FILLING: 0:59
HEATING: 30:34
STERILIZING: 4:00
VENTING: 3:13
DRYING: 28:51
TOTAL CYCLE: 01:07:37

06/14/2021 08:38 AM
ACCEPTED: X
REJECTED:
Operator: U01