
Midmark M11 - v1.0.5
Total Cycles: 3227

Name: RITTER2
Sterilizer ID: V2313228
12/18/2023 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.2 F 0.0
2:00 153.7 F 0.9
4:00 158.5 F 3.9
6:00 182.1 F 7.9
8:00 202.1 F 12.7
10:00 213.6 F 7.1
12:00 228.1 F 12.2
14:00 234.1 F 9.1
16:00 237.7 F 10.0
18:00 245.6 F 13.7
20:00 257.2 F 20.0
22:00 268.1 F 26.7
24:00 271.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.7
1:00 272.7 F 30.1
2:00 271.6 F 29.8
3:00 272.6 F 30.1
4:00 271.5 F 29.8

Min 271.4 F 29.5
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.0 F 0.0
5:00 224.6 F 0.0
10:00 199.9 F 0.0
15:00 186.1 F 0.0
20:00 180.3 F 0.0
25:00 178.9 F 0.0
30:00 178.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:06
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:02:22
CYCLE COMPLETE

12/18/2023 11:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____