

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000197  
8:00 21/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:10  
PRESSURIZING 3:18  
STERILIZING 4:53  
136.0 C 229kPa 4:53  
136.4 C 233kPa 5:23  
136.4 C 231kPa 5:53  
136.4 C 233kPa 6:23  
136.3 C 232kPa 6:53  
136.3 C 232kPa 7:23  
136.4 C 231kPa 7:53  
136.3 C 232kPa 8:23  
136.3 C 232kPa 8:23  
VENTING 8:23  
AIR DRYING 9:22  
CYCLE COMPLETE 11:45

-----  
Digital Signature #  
8431805200E00667  
-----