
Midmark M11 - v1.0.0
Total Cycles: 3419

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2023 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.2 F 0.0
2:00 188.8 F 0.9
4:00 187.5 F 4.4
6:00 203.7 F 9.0
8:00 222.2 F 8.4
10:00 229.5 F 9.1
12:00 237.2 F 10.1
14:00 242.1 F 12.0
16:00 240.6 F 10.5
18:00 254.1 F 17.2
20:00 265.6 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.6 F 28.9
2:00 271.9 F 28.2
3:00 271.5 F 28.3
4:00 272.6 F 28.9

Min 271.3 F 27.9
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 213.1 F 0.0
10:00 230.8 F 0.0
15:00 224.2 F 0.0
20:00 226.4 F 0.0
25:00 223.3 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

06/21/2023 09:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____