
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 004533
8:08 17/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:33
STERILIZING 3:37
136.5 C 222kPa 3:37
136.1 C 222kPa 4:07
135.8 C 220kPa 4:37
136.2 C 222kPa 5:07
136.3 C 223kPa 5:37
136.4 C 223kPa 6:07
136.4 C 224kPa 6:37
136.4 C 220kPa 7:07
136.4 C 220kPa 7:07
VENTING 7:08
AIR DRYING 7:55
CYCLE COMPLETE 8:27

Digital Signature #
9363A970A3A9D6A0
