
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001543
9:31 10/29/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:38
PRESSURIZING 6:52
STERILIZING 8:36
136.3 C 221kPa 8:36
136.3 C 221kPa 9:06
136.2 C 218kPa 9:36
136.3 C 221kPa 10:06
136.1 C 219kPa 10:36
136.1 C 219kPa 11:06
136.4 C 222kPa 11:37
136.0 C 220kPa 12:06
VENTING 12:06
AIR DRYING 13:11
CYCLE COMPLETE 43:11

Digital Signature #
BF580A116D64F8D5
