
Midmark M11 - v1.0.5
Total Cycles: 510

Name: RITTER2
Sterilizer ID: V2313228
11/01/2021 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.7 F 0.0
2:00 178.5 F 3.0
4:00 194.0 F 7.9
6:00 214.1 F 13.0
8:00 229.2 F 8.1
10:00 239.8 F 13.3
12:00 240.2 F 11.4
14:00 239.5 F 10.5
16:00 253.0 F 17.2
18:00 264.4 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.5 F 29.4
2:00 271.3 F 28.8
3:00 272.5 F 29.5
4:00 271.5 F 28.9

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 236.8 F 0.0
10:00 226.1 F 0.0
15:00 184.5 F 0.0
20:00 179.7 F 0.0
25:00 177.2 F 0.0
30:00 176.5 F 0.0

Drying Complete

FILLING: 2:02
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

11/01/2021 12:16 PM
ACCEPTED: X
REJECTED:
Operator: U01