
Midmark M11 - v1.0.5
Total Cycles: 4286

Name: RITTER2
Sterilizer ID: V2313228
01/23/2025 11:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.1 F 0.0
2:00 177.0 F 1.6
4:00 177.6 F 5.2
6:00 199.1 F 9.8
8:00 224.3 F 7.2
10:00 228.6 F 8.1
12:00 236.7 F 12.7
14:00 236.0 F 9.6
16:00 239.1 F 9.9
18:00 246.4 F 13.9
20:00 257.9 F 20.0
22:00 268.0 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.4 F 29.6
2:00 271.6 F 28.9
3:00 272.4 F 29.9
4:00 272.1 F 29.3

Min 271.0 F 28.7
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.6 F 0.0
5:00 235.9 F 0.0
10:00 222.2 F 0.0
15:00 202.4 F 0.0
20:00 187.8 F 0.0
25:00 180.3 F 0.0
30:00 199.0 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:02:26
CYCLE COMPLETE

01/23/2025 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____