
Midmark M11 - v1.0.0
Total Cycles: 4386

Name: RITTER 1
Sterilizer ID: V2178198
03/20/2024 12:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.3 F 0.0
2:00 114.8 F 0.6
4:00 122.1 F 2.7
6:00 134.7 F 5.5
8:00 161.2 F 9.0
10:00 184.3 F 13.1
12:00 207.0 F 4.8
14:00 219.6 F 8.6
16:00 231.9 F 12.8
18:00 233.6 F 9.0
20:00 241.3 F 11.6
22:00 243.0 F 12.3
24:00 254.1 F 18.1
26:00 264.6 F 24.3
28:00 271.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.4
2:00 271.6 F 29.3
3:00 272.4 F 29.4
4:00 271.4 F 29.2

Min 271.3 F 28.9
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 223.5 F 0.0
10:00 232.1 F 0.0
15:00 234.2 F 0.0
20:00 227.1 F 0.0
25:00 226.5 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:02
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:06:47
CYCLE COMPLETE

03/20/2024 01:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____