
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 003616
13:36 05/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:42
STERILIZING 3:53
136.4 C 220kPa 3:53
136.2 C 218kPa 4:23
136.5 C 225kPa 4:53
136.4 C 225kPa 5:23
136.3 C 224kPa 5:53
136.4 C 222kPa 6:23
136.3 C 224kPa 6:53
136.1 C 220kPa 7:23
136.1 C 220kPa 7:23
VENTING 7:23
AIR DRYING 8:10
CYCLE COMPLETE 8:30

Digital Signature #
1148F1D43A09E406
