
Midmark M11 - v1.0.5
Total Cycles: 3850

Name: RITTER2
Sterilizer ID: V2313228
08/12/2024 04:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.0 F 0.0
2:00 179.8 F 3.3
4:00 195.9 F 8.2
6:00 216.3 F 13.2
8:00 225.8 F 8.5
10:00 235.6 F 11.6
12:00 241.0 F 12.0
14:00 240.6 F 11.2
16:00 253.5 F 17.7
18:00 265.1 F 24.5
20:00 271.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.4 F 29.7
2:00 272.3 F 30.0
3:00 272.2 F 29.5
4:00 272.7 F 30.2

Min 271.6 F 29.3
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 221.7 F 0.0
10:00 227.8 F 0.0
15:00 224.4 F 0.0
20:00 221.2 F 0.0
25:00 222.2 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 20:33
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:00:10
CYCLE COMPLETE

08/12/2024 05:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____