
Midmark M11 - v1.0.0
Total Cycles: 1853

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2022 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.8 F 0.0
2:00 179.1 F 3.5
4:00 199.9 F 8.0
6:00 218.2 F 13.3
8:00 227.3 F 7.7
10:00 240.1 F 12.8
12:00 238.0 F 9.9
14:00 237.3 F 8.6
16:00 249.5 F 14.7
18:00 261.4 F 21.5
20:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.8 F 28.8
2:00 271.8 F 28.1
3:00 272.0 F 28.6
4:00 272.4 F 28.6

Min 271.2 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.7 F 0.0
5:00 239.7 F 0.0
10:00 169.2 F 0.0
15:00 160.6 F 0.0
20:00 181.1 F 0.0
25:00 160.5 F 0.0
Dry Cycle Aborted

FILLING: 1:45
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:10
DRYING: 29:41
TOTAL CYCLE: 00:59:07

04/21/2022 11:04 AM
ACCEPTED: X
REJECTED:
Operator: U02