
Midmark M11 - v1.0.5
Total Cycles: 3573

Name: RITTER2
Sterilizer ID: V2313228
05/01/2024 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.1 F 0.0
2:00 76.9 F 0.1
4:00 96.5 F 1.5
6:00 133.0 F 4.5
8:00 165.7 F 7.8
10:00 192.3 F 11.4
12:00 208.1 F 3.8
14:00 216.0 F 7.5
16:00 227.8 F 11.6
18:00 229.2 F 7.4
20:00 241.7 F 12.8
22:00 240.5 F 11.1
24:00 253.0 F 17.3
26:00 264.0 F 23.8
28:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.5
2:00 271.4 F 29.1
3:00 272.7 F 29.6
4:00 271.5 F 28.8

Min 271.4 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.5 F 0.0
5:00 199.9 F 0.0
10:00 194.1 F 0.0
15:00 187.0 F 0.0
20:00 187.3 F 0.0
25:00 194.6 F 0.0
30:00 199.2 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 28:06
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:05:55
CYCLE COMPLETE

05/01/2024 09:44 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____